



LAWRENCE
LIVERMORE
NATIONAL
LABORATORY

Preconcentration of cesium-137 (^{137}Cs) from large volume water samples using zirconium ferrocyanide embedded on cartridge water filters

Terry Hamilton, Roger Martinelli, Steven Kehl and Jim Brunk

Lawrence Livermore National Laboratory
PO Box 808, Livermore, CA 94551-0808, U.S.A.

November 2005

For submission to:
Methods and Applications of Radioanalytical Chemistry
MARC VII
Kailua-Kona, Hawaii, USA
April 3-7 2006

DISCLAIMER

This document was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor the University of California nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or the University of California, and shall not be used for advertising or product endorsement purposes.

Work performed under the auspices of the U.S. Department of Energy by the University of California, Lawrence Livermore National Laboratory under Contract No. W-7405-Eng-48.

Preconcentration of cesium-137 (^{137}Cs) from large volume water samples using zirconium ferrocyanide embedded on cartridge water filters.

Terry Hamilton, Roger Martinelli, Steven Kehl and Jim Brunk

Lawrence Livermore National Laboratory
PO Box 808, Livermore, CA 94551-0808, U.S.A.

The determination of cesium-137 (^{137}Cs) in surface waters usually requires some form of preconcentration prior to radiometric analysis. There are many types of inorganic sorbents suitable for this application including ammonium molybdophosphates (AMP), natural and synthetic zeolites, and ferrocyanides of transition metals. Ferrocyanides are relatively easy to prepare and generally provide high selectivity for separation of cesium in the presence of alkaline and alkali metals. This paper describes a method for preparing a zirconium ferrocyanide (ZrFeCN) based sorbent material that is conveniently embedded on a polyethylene wound water filter cartridge (0.5 μm Micro-Wynd®, Cuno Inc.). These ZrFeCN based sorbent cartridges have been used extensively at former nuclear test sites in the Marshall Islands for preconcentration of ^{137}Cs in water samples collected on plate lysimeters. At pump rates of < 2 liters per minute, the ZrFeCN cartridge collections exhibit a very high efficiency for cesium uptake (>95%) and good stability for long-term, automated field applications. The use of a synthetic support base (in this case, polyethylene) for the sorbent was preferred over more commonly used natural cotton materials because of its lower ^{137}Cs background and ash content.

Work performed under the auspices of the U.S. Department of Energy by the University of California, Lawrence Livermore National Laboratory under Contract No. W-7405-Eng-48.

Methods and Applications of Radioanalytical Chemistry
MARC VII
Kailua-Kona, Hawaii, USA
April 3-7 2006

University of California
Lawrence Livermore National Laboratory
Technical Information Department
Livermore, CA 94551

